



HAL
open science

Polyelectrolytes Multilayer Films on titanium and titanium alloys for biomedical applications

C Brunot, L Mora-Ponsonnet, D Décoret, C Picart, B Beaugiraud, Brigitte Grosgeat

► **To cite this version:**

C Brunot, L Mora-Ponsonnet, D Décoret, C Picart, B Beaugiraud, et al.. Polyelectrolytes Multilayer Films on titanium and titanium alloys for biomedical applications. *eCells and Materials Journal*, 2014, eCM-2014, 27 (suppl.1), pp.17. hal-02004864

HAL Id: hal-02004864

<https://hal.univ-reims.fr/hal-02004864>

Submitted on 2 Feb 2019

HAL is a multi-disciplinary open access archive for the deposit and dissemination of scientific research documents, whether they are published or not. The documents may come from teaching and research institutions in France or abroad, or from public or private research centers.

L'archive ouverte pluridisciplinaire **HAL**, est destinée au dépôt et à la diffusion de documents scientifiques de niveau recherche, publiés ou non, émanant des établissements d'enseignement et de recherche français ou étrangers, des laboratoires publics ou privés.

Polyelectrolytes Multilayer Films on titanium and titanium alloys for biomedical applications

C. Brunot^{1,2}, L. Mora-Ponsonnet³, D. Décoret¹, C. Picart⁴, B. Beaugiraud⁵, et B. Grosogeat^{1,6}

¹Laboratoire des Multimatériaux et Interfaces, UMR CNRS 5615, Université de Lyon, Equipe Biomateriaux et Interfaces Biologiques, UFR d'Odontologie, rue Guillaume Paradin, 69372 Lyon Cedex 08. ²Faculté d'Odontologie, Université de Reims Champagne-Ardenne, 2 rue du Général Koenig, 51100 Reims. ³Laboratoire de Bio-Ingénierie de Polymères Cardiovasculaires, INSERM U 698, Institut Galilée, Université Paris 13, 99 avenue JB Clément, 93430 Villetaneuse. ⁴Laboratoire des Matériaux et du Génie Physique, UMR CNRS 5628, MINATEC, 3 parvis Louis Néel, 38016 Grenoble. ⁵Laboratoire de Tribologie et Dynamique des Systèmes, Ecole Centrale de Lyon, 36 avenue Guy de Collongue, 69134 Ecully Cedex. ⁶Hospices Civils de Lyon, Service de Consultations et de Traitements Dentaires, 6-8 Place Depéret, 69365 Lyon Cedex 07.

INTRODUCTION: The aim of this research was to optimize titanium (Ti) and titanium alloy (NiTi) surfaces coated with Polyelectrolytes Multilayer Films (PMF)^[1]. We studied different biomedical parameters related to this surface treatment in order to develop specific biomedical applications in the dental field (dental implants, endodontic instruments, orthodontic arches). Firstly, we determined if PMF had a detectable physisorption on Ti and NiTi. Secondly, we studied biocompatibility of poly-ethyleneimine (PEI), the precursor based-layer of PMF. Finally, sterilization tests were realized to analyze their potential impact on the physico-chemical structure of PMF.

METHODS: Two types of films were characterized : polystyrene-sulfonate/polyallylamine hydrochloride films PEI-(PSS/PAH) and hyaluronic acid/poly-L-lysine films PEI-(HA/PLL). Physico-chemical characterization was carried out by tensiometry, atomic force microscopy (AFM), and confocal microscopy. A biological study using human fibroblasts was carried out. Cell response was observed after 0, 2, 4 and 7 days *in vitro* using morphologic criteria (scanning electron microscopy), adhesion (fluorescence microscopy image analysis), and proliferation (Methyl Tetrazolium Test). Three methods of PMF sterilization were compared: ultraviolets (UV), ethylene oxide (ETO), and autoclave. PSS/PAH films were characterized by the same physico-chemical tests and biological studies after the sterilization process.

RESULTS: Results showed that PSS/PAH films were more biocompatible than HA/PLL films on both metallic biomaterials. The precursor based-layer study demonstrated that

PEI was not biocompatible^[3]. PMF sterilized by autoclave showed similar biocompatibility. Such results could not be found with UV and ETO. Nevertheless, some investigations have to be realized to prove that the structure of PMF was not perturbed after these processes.

DISCUSSION & CONCLUSION: Based on our results, we decided to use PAH as the precursor based-layer (positively charged) and to stop using PEI on biomaterials made of Ti or NiTi. We are now planning to integrate bioactive molecules between layers of the PMF. To this aim other parameters need to be studied to characterize *in vivo* potential biomedical application, such as films aging in the oral cavity, as well as in salivary or fluoride environment.

REFERENCES:

- ^[1]Decher G. Fuzzy nanoassemblies: toward layered polymeric multicomposites. *Science* 1997,277:1232-7. ^[2]Brunot, C, Grosogeat B, Picart C, Lagneau C, Jaffrezic-Renault N, Ponsonnet L. Response of fibroblast activity and polyelectrolyte multilayer films coating titanium. *Dent Mater* 2008,24(8):1025-35. ^[3]Brunot C, Ponsonnet L, Lagneau C, Farge P, Picart C, Grosogeat B. Cytotoxicity of polyethyleneimine (PEI), precursor base layer of polyelectrolyte multilayers films. *Biomaterials* 2007,28(4):632-40.

ACKNOWLEDGEMENTS: Timet[®]Savoie society (Ugine, France) and AMF[®] society (Lury sur Arnon, France) for providing samples.