



**HAL**  
open science

## **Matrix-assisted laser desorption ionization-time of flight identification of *Schizophyllum commune*: perspectives on the review by Chowdhary et al.**

Antoine Huguenin, Aurélie Lorot, Dominique Zachar, Oriane Dudez, Alexandre Mzabi, Emilie Dupuis, Véronique Lehrter, Jean-Claude Merol, Isabelle Villena, Jérôme Depaquit, et al.

### ► To cite this version:

Antoine Huguenin, Aurélie Lorot, Dominique Zachar, Oriane Dudez, Alexandre Mzabi, et al.. Matrix-assisted laser desorption ionization-time of flight identification of *Schizophyllum commune*: perspectives on the review by Chowdhary et al.. *Medical Mycology*, 2015, 53 (8), pp.896 - 897. 10.1093/mmy/myv031 . hal-03045895

**HAL Id: hal-03045895**

**<https://hal.univ-reims.fr/hal-03045895>**

Submitted on 8 Dec 2020

**HAL** is a multi-disciplinary open access archive for the deposit and dissemination of scientific research documents, whether they are published or not. The documents may come from teaching and research institutions in France or abroad, or from public or private research centers.

L'archive ouverte pluridisciplinaire **HAL**, est destinée au dépôt et à la diffusion de documents scientifiques de niveau recherche, publiés ou non, émanant des établissements d'enseignement et de recherche français ou étrangers, des laboratoires publics ou privés.



## Letter to the Editors

# Matrix-assisted laser desorption ionization-time of flight identification of *Schizophyllum commune*: perspectives on the review by Chowdhary *et al.*

Antoine Huguenin<sup>1</sup>, Aurélie Lorot<sup>2</sup>, Dominique Zachar<sup>3</sup>, Oriane Dudez<sup>3</sup>, Alexandre Mzabi<sup>1,5</sup>, Emilie Dupuis<sup>1,5</sup>, Véronique Lehrter<sup>4</sup>, Jean-Claude Merol<sup>2</sup>, Isabelle Villena<sup>1,5</sup>, Jérôme Depaquit<sup>1,4</sup> and Dominique Toubas<sup>1,6,\*</sup>

<sup>1</sup>Laboratoire de Parasitologie Mycologie, CHU de Reims, Université Reims Champagne Ardenne, Hôpital Maison Blanche, 45 rue Cognacq Jay, 51092 Reims cedex, France, <sup>2</sup>Service d'ORL et de Chirurgie Cervico-Faciale, CHU de Reims, Hopital Robert Debré, 45 rue Cognacq Jay, 51092 Reims cedex, France, <sup>3</sup>Laboratoire central d'Anatomie et de Cytologie pathologiques, CHU de Reims, Hôpital Robert-Debré, Université Reims Champagne-Ardenne, rue du Général-Koenig, 51100 Reims cedex, France, <sup>4</sup>Université Reims Champagne-Ardenne, ANSES, EA4688 - USC 'transmission vectorielle et épidémiologie de maladies parasitaires (VECPAR)', SFR Cap Santé, Faculté de Pharmacie, 51 rue Cognacq-Jay, 51096 Reims, France, <sup>5</sup>EA 3800, Université Reims Champagne-Ardenne, France and <sup>6</sup>Equipe MéDIAN, Unité MEDyC CNRS UMR 7369, Université de Reims Champagne-Ardenne, France

\*To whom correspondence should be addressed. Dr Dominique Toubas, Laboratoire de Parasitologie Mycologie, CHU de Reims, Hôpital Maison Blanche, 45 rue Cognacq Jay, 51092 Reims cedex, France. E-mail: [dtoubas@chu-reims.fr](mailto:dtoubas@chu-reims.fr)

Received 23 January 2015; Revised 27 March 2015; Accepted 6 April 2015

*SIR*: The excellent review of Chowdhary *et al.*<sup>1</sup> emphasizes the significance of filamentous basidiomycetes (FBM) in human pathology. This review raises the challenge of conventional identification of *Schizophyllum commune* and other FBM. Due to the absence of asexual reproduction, monokaryotic isolates usually grow as nonsporulating molds (NSM) lacking characteristic morphological features.<sup>2,3</sup> In dikaryotic strains, those features (clamp connections or spicules) can be lacking or take several weeks to appear. Moreover they are insufficient to identify the species. Induction of characteristic fruiting bodies is inconstant and usually requires 3 weeks.<sup>4</sup> FBM identification is therefore based on molecular analysis.<sup>5</sup> Since the introduction of molecular identification in medical mycology, the number of documented cases involving *S. commune* and

other FBM has dramatically increased.<sup>1,6</sup> In a retrospective molecular analysis of 52 NSM strains isolated from respiratory tract specimens,<sup>4</sup> FBM represented 96% of the isolates including *S. commune* (52%). Despite their important contribution, molecular techniques have several limitations. They are time consuming and expensive and interpretation could be difficult.<sup>7</sup> We would like to complete these two identification approaches presented by Chowdhary *et al.* by matrix-assisted laser desorption ionization time-of-flight mass spectrometry (MALDI-TOF MS).

MALDI-TOF MS is a revolutionary tool in clinical microbiology. It is a soft-ionization mass spectrometry technique based on co-crystallization of the sample with an absorbent matrix allowing acquisition of protein mass spectra "fingerprints." Comparison of the mass spectral profiles

(MSP) to a database allows a quick and easy identification of microorganisms. The equipment is expensive, but identification of a strain is particularly rapid and cost-effective (less than \$1 by identification<sup>8</sup>). This technique is increasingly used in routine to identify bacteria and yeasts at species level in a few minutes.<sup>9,10</sup> More and more studies focus on identification of filamentous fungi.<sup>11–13</sup> Therefore, MALDI-TOF MS presents a great potential for identification of strains without fructifications, especially NSM identification. Identification of environmental FBM from rotten wood by MALDI-TOF MS has been described in 2005.<sup>14</sup> More recently, identification of several FBM strains isolated from clinical samples by MALDI-TOF MS has been reported.<sup>13,15</sup>

In our lab, we systematically analyze with MALDI-TOF MS, every NSM isolated from clinical samples. We have recently identified a strain of *S. commune* isolated from a sinusitis case [A. Huguenin, unpublished results]. In our laboratory, it is the first time we identify a strain of *S. commune* isolated from a clinical sample. MALDI-TOF MS has permitted identification at species level of this strain 9 days before we could observe characteristic features of FBM (spicules and clamp-connections). Although there is currently a lack of published reports on this subject, similar results have been observed in other clinical laboratories [J. Michel, personal communication; B. Sendid, personal communication].

MALDI TOF is however limited by the lack of diversity among MSP in the provider's MALDI TOF library: for example *S. commune* is the only species of FBM represented in the Bruker Fungi Library V1.0. The NIH library including 11 MSP of FBM, has permitted to identify only 12 (32%) of 37 FBM clinical isolates.<sup>15</sup> Many species considered as nonpathogenic, are not represented in the library. This issue could be addressed by the development of an in-house library for MSP. Moreover, the quality of MSP could be an important issue. Phenotypic state (conidia, basidiocarp) may be a variability source of measured spectra, making essential to define standardized conditions for the production of MSP and the implementation of the spectral data base.

In conclusion, *S. commune* and other FBM are emerging agents in human diseases. MALDI-TOF MS has an important potential for a fast and inexpensive identification of *S. commune* and other FBM. It could contribute to improve epidemiological data of these infections. For this purpose, improvement in reference spectra library, both in diversity and quality, is a must.

We thank Dr. P.L. Toubas for kindly reviewing the manuscript and Dr. L Brasme for his help with the MALDI-TOF analysis technics.

## References

1. Chowdhary A, Kathuria S, Agarwal K et al. Recognizing filamentous basidiomycetes as agents of human disease: a review. *Med Mycol* 2014; 52: 782–797.
2. Sigler L, la Maza de LM, Tan G et al. Diagnostic difficulties caused by a nonclamped *Schizophyllum commune* isolate in a case of fungus ball of the lung. *J Clin Microbiol* 1995; 33: 1979–1983.
3. Gari-Toussaint M, Lachaud L, Pihet M et al. Sinusite récidivante due à *Schizophyllum commune*. À propos de deux nouveaux cas. *J Mycol Med* 2010; 21: 289–293.
4. Singh PK, Kathuria S, Agarwal K et al. Clinical significance and molecular characterization of nonsporulating molds isolated from the respiratory tracts of bronchopulmonary mycosis patients with special reference to basidiomycetes. *J Clin Microbiol* 2013; 51: 3331–3337.
5. de Hoog GSG, van den Ende AHAG. Molecular diagnostics of clinical strains of filamentous basidiomycetes. *Mykosen* 1998; 41: 183–189.
6. Chowdhary A, Randhawa HS, Gaur SN et al. *Schizophyllum commune* as an emerging fungal pathogen: a review and report of two cases. *Mycoses* 2012; 56: 1–10.
7. Romanelli AM, Sutton DA, Thompson EH et al. Sequence-based identification of filamentous basidiomycetous fungi from clinical specimens: a cautionary note. *J Clin Microbiol* 2010; 48: 741–752.
8. Gaillot O, Blondiaux N, Loïez C et al. Cost-effectiveness of switch to matrix-assisted laser desorption ionization-time of flight mass spectrometry for routine bacterial identification. *J Clin Microbiol* 2011; 49: 4412.
9. Lavigne J-P, Espinal P, Dunyach-Remy C et al. Mass spectrometry: a revolution in clinical microbiology? *Clin Chem Lab Med* 2013; 51: 257–270.
10. Bizzini A, Greub G. Matrix-assisted laser desorption ionization time-of-flight mass spectrometry, a revolution in clinical microbial identification. *Clin Microbiol Infect* 2010; 16: 1614–1619.
11. De Carolis E, Posteraro B, Lass-Flörl C et al. Species identification of *Aspergillus*, *Fusarium* and *Mucorales* with direct surface analysis by matrix-assisted laser desorption ionization time-of-flight mass spectrometry. *Clin Microbiol Infect* 2011; 18: 475–484.
12. Becker PT, de Bel A, Martiny D et al. Identification of filamentous fungi isolates by MALDI-TOF mass spectrometry: clinical evaluation of an extended reference spectra library. *Med Mycol* 2014; 52: 826–834.
13. Gautier M, Ranque S, Normand A-C et al. MALDI-TOF mass spectrometry: revolutionising clinical laboratory diagnosis of mould infections. *Clin Microbiol Infect* 2014; 20: 1366–1371.
14. Schmidt O, Kallow W. Differentiation of indoor wood decay fungi with MALDI-TOF mass spectrometry. *Holzforschung* 2005; 3: 374–377.
15. Lau AF, Drake SK, Calhoun LB et al. Development of a clinically comprehensive database and a simple procedure for identification of molds from solid media by matrix-assisted laser desorption ionization-time of flight mass spectrometry. *J Clin Microbiol* 2013; 51: 828–834.